

## Review Article

# The Use of Highly Sensitive Detection Methods for Eradication of *Plasmodium*

Marianne Smedegaard Hede<sup>1</sup> and Birgitta Ruth Knudsen<sup>2\*</sup><sup>1</sup>Zymonostics ApS, Denmark<sup>2</sup>Department of Molecular Biology and Genetics, University of Aarhus, Denmark

## \*Corresponding author

Birgitta Ruth Knudsen, Department of Molecular Biology and Genetics, University of Aarhus, CF Møllers Alle 3, Build. 1130 Room 311, 8000 Aarhus C, Denmark, Tel: 60202673; Fax: 86196500; Email: brk@mbg.au.dk

Submitted: 26 September 2017

Accepted: 20 October 2017

Published: 23 October 2017

ISSN: 2475-9430

## Copyright

© 2017 Hede et al.

OPEN ACCESS

## Abstract

The key to a successful malaria eradication program is highly efficient detection of *Plasmodium* infected people followed by appropriate treatment to avoid spreading of the parasite. We will discuss some of the demands that such a detection method needs to fulfill and review some of the advantages and disadvantages of currently available detection methods.

## Keywords

- *Plasmodium*
- Malaria elimination
- Nucleic acid amplification techniques

## ABBREVIATIONS

LAMP: Loop Mediated Isothermal Amplification; NAAT: Nucleic Acid Amplification Techniques; PCR: Polymerase Chain Reaction; Ptopoi: *Plasmodium* topoisomerase I; REEAD: Rolling Circle-Enhanced Enzyme Activity Detection

## INTRODUCTION

An international effort against malaria has managed to reduce the number of deaths caused by malaria to 30% in a five-year period from 2010 to 2015 [1]. Despite this promising development, almost half a million people still die from malaria every year. Most of these are children below 5 years of age [1]. In Africa alone, the economic burden of malaria is estimated to be around 12 billion USD in lost gross domestic product. The serious consequences of such loss for society as well as for the individual are enhanced by malaria being predominant in some of the poorest countries in the world [2,3]. It is therefore clear that a continued and optimized strategy to fight malaria is highly needed. Malaria is an infectious disease caused by infection with a parasite of the *Plasmodium* genus. Five members of this genus (*P. falciparum*, *P. vivax*, *P. ovale*, *P. malariae* and *P. knowlesi*) are known to cause malaria in humans. All five species are transferred by mosquitoes belonging to the *Anopheles* genus which is the host for part of the parasitic life cycle [4]. In areas where malaria is endemic, a large fraction (up to 40%) of the population will carry the *Plasmodium* parasite in their blood even if they are not displaying symptoms of malaria [5-8]. In fact, 92% of the *Plasmodium* infections found in a Kenyan screening study were asymptomatic [7]. The high prevalence of asymptomatic *Plasmodium* infections is true even for areas with a low *Plasmodium* burden [6]. Therefore, for eradication purposes very sensitive methods for detection of *Plasmodium* infections are required for finding and treating

people unknowingly infected with the *Plasmodium* parasite. Apart from sensitivity, there are also a number of other demands which should be met by the optimal detection methodology. In surveillance projects a large number of asymptomatic individuals may be screened. The ideal method for use in surveillance and eradication should therefore be cost efficient, easy to use, and suitable for high throughput screening and, preferably, allow the use of non-invasive sample types such as saliva.

## CURRENT METHODS FOR DETECTION OF PLASMODIUM INFECTIONS

### Blood smear microscopy

The gold standard within malaria detection is blood smear microscopy [9]. While this method carries the benefit of direct observation of the *Plasmodium* parasite, microscopy is not suitable for high throughput screening and is relatively time consuming. Moreover, both specificity and sensitivity will depend on the experience of the microscopist. The detection level in a typical setting is 50-100 parasites per  $\mu\text{L}$  blood [10] which is sufficient for diagnosis in most cases but not necessary low enough to facilitate eradication of the *Plasmodium* parasite.

### Rapid diagnostic tests

Antibody based rapid diagnostic tests (RDTs) have gained much popularity for diagnosis in endemic areas [11]. RDTs are easy to use and can be used with limited or no training. Moreover, they carry the advantage of being usable in low resource settings and as point-of care tests [11-13]. Unfortunately, the detection limits of RDTs are relatively high and RDTs are therefore not optimal for eradication studies as they show limited sensitivity for samples with low parasitemia [11]. Furthermore, mutations in the biomarker have been reported to lead to false negatives [14,15].

## PCR

In recent years methods for detection of *Plasmodium* infections using nucleic acid amplification techniques (NAAT) have emerged [16]. The best established of these are PCR (polymerase chain reaction). PCR is an extremely sensitive method for detection of specific DNA sequences. When used for malaria diagnosis a part of the *Plasmodium* genome such as the 18S small-subunit RNA genes is used as a template and detected by PCR amplification [17,18]. The PCR protocols may be designed to detect only a single *Plasmodium* species or it may be designed to detect all members of the genus depending on the needs in a given situation [19, 20]. Most protocols use blood as sample type but also non-invasive sample types such as saliva and urine has been used. As mentioned, such non-invasive sample types can be used with benefit if a high number of asymptomatic persons are to be tested as part of a surveillance study [21-23]. Theoretically, as little as a single target DNA molecule can be detected using PCR making it an interesting methodology for eradication studies. In concert with the unsurpassable theoretical sensitivity of PCR, PCR protocols for detection of *Plasmodium* parasites has generally reported very high sensitivities and detection limits well below 1 parasite per  $\mu\text{L}$  blood [16,24,25]. Unfortunately, the ability of detecting a few target molecules and the exponential nature of PCR amplification also inevitably leads to a risk of false positives and reduced specificity. Another inherent risk is false negatives due to mutations in the *Plasmodium* target DNA. Indeed, as the burden of *Plasmodium* infection in a region is lowered and the attempts to eradicate the parasite are intensified, there will be a greatly increased evolutionary pressure on the *Plasmodium* parasite to mutate in order to render itself undetectable for the commonly used detection methods. Finally, it should be mentioned that PCR sample preparation and analysis requires trained personnel as well as specialized and expensive equipment, which may not be available to local laboratories in the relevant countries.

## LAMP

Another NAAT method that has attracted much attention is LAMP (Loop Mediated isothermal amplification) [16,26,27]. As in the case of PCR the LAMP technology depends on primer based DNA target amplification. This gives LAMP the same advantages as PCR when it comes to flexibility in assay design and allow either specie specific or genus specific detection of *Plasmodium* [28-30]. However, the risk of mutations giving rise to false negatives is also a problem for LAMP. In contrast to PCR, LAMP benefits from the amplification step being isothermal and the technique is hence more usable for low resource settings than PCR. The detection limit of LAMP protocols are down to 1-2 parasite per  $\mu\text{L}$  blood and when compared to PCR, the sensitivity often exceeds 95% [16,30,31]. The detection limit of LAMP is thus better than for blood smear microscopy but may still not be good enough for eradication studies.

## REEAD

A fundamentally different nucleic acid amplification based detection method, REEAD (rolling circle enhanced enzyme activity detection), has recently been described [32,33]. As

implied by the name REEAD allows the detection of *Plasmodium* through detection of the activity of a *Plasmodium* encoded enzyme, topoisomerase I (pTopoI). The core of the REEAD technology is pTopoI mediated conversion of a specific DNA substrate into a DNA circle which is amplified using rolling circle amplification and detected [32,33]. Topoisomerases are found in high copy numbers in all living cells from bacteria to humans and their activity is essential to cell survival [34]. When using pTopoI activity as biomarker, the risk of getting false negatives due to mutations is therefore very low. Neither the pTopoI reaction nor the rolling circle amplification step used for amplification of the pTopoI generated signals require heating or specialized equipment. Furthermore, the read-out, which is currently based on fluorescence microscopy, can easily be adapted for user-friendly and high throughput use. Such adaptation can also render the REEAD assay well suited for use in low-resource settings and even for use completely without electricity by exploding e.g. a horseradish peroxidase mediated color reaction for visualization (Hede et al. unpublished data). The detection limit of REEAD was found to be as low as 0.06 parasites per  $\mu\text{L}$  blood. Together with the fact that REEAD allows detection of the *Plasmodium* parasite using saliva as test material the low detection limit makes it a very promising emerging technology for malaria eradication purposes [33].

## DISCUSSION AND CONCLUSION

In conclusion, malaria elimination projects present a unique set of challenges to detection methods. First of all, the detection method should detect the presence of *Plasmodium* parasite directly, it should be suitable for high throughput screenings, and be sensitive enough to detect even very low concentrations of *Plasmodium*. Secondly the method should be easy to operate in a low technological setting and not be susceptible to false negatives due to mutations in the biomarker. Lastly the detection method should ideally allow testing of non-invasive sample sets such as urea or saliva. As summarized in Table 1, none of the currently used malaria diagnostic tests meet all these criteria. Microscopy, RDT's, PCR, and LAMP all face inherent issues in

**Table 1:** Comparison of malaria detection methods - indicates "not possible", + indicates "possible"; (+) indicates potentially possible.

	Microscopy	RDT	PCR	LAMP	REEAD
Adaptable for high throughput screening	-	+	+	+	+
Detection limit below 1 parasite/ $\mu\text{L}$	-	-	+	(+)	+
Adaptable for use without electricity	-	+	-	(+)	+
Low cost	+	(+)	+	+	+
Vulnerable to mutations	-	+	+	+	-
non-invasive sample types	-	-	+	+	+

terms of sensitivity or liability to mutations that will be difficult if not impossible to circumvent. Hopefully, further developments on NAAT techniques based on detection of essential enzymatic activities such as REEAD will be able to circumvent these difficulties and pave the way for better, more efficient eradication programs.

## REFERENCES

1. <http://www.who.int/malaria/media/world-malaria-report-2016/en/>
2. Gallup JL, Sachs JD. The economic burden of malaria. *The Am J Trop Med Hyg.* 2001; 64: 85-96.
3. [https://www.cdc.gov/malaria/malaria\\_worldwide/impact.html](https://www.cdc.gov/malaria/malaria_worldwide/impact.html)
4. Meibalan E, Marti M. *Biology of Malaria Transmission.* Cold Spring Harb Perspect Med. 2017; 7.
5. Golassa L, Enweji N, Erko B, Aseffa A, Swedberg G. Detection of a substantial number of sub-microscopic *Plasmodium falciparum* infections by polymerase chain reaction: a potential threat to malaria control and diagnosis in Ethiopia. *Malar J.* 2013; 12: 352.
6. Saenz FE, Arevalo-Cortes A, Valenzuela G, Vallejo AF, Castellanos A, Poveda-Loayza AC, et al. Malaria epidemiology in low-endemicity areas of the northern coast of Ecuador: high prevalence of asymptomatic infections. *Malar J.* 2017; 16: 300.
7. Idris ZM, Chan CW, Kongere J, Gitaka J, Logedi J, Omar A, et al. High and Heterogeneous Prevalence of Asymptomatic and Sub-microscopic Malaria Infections on Islands in Lake Victoria, Kenya. *Sci Rep.* 2016; 6.
8. Elbadry MA, Al-Khedery B, Tagliamonte MS, Yowell CA, Raccurt CP, Existe A, et al. High prevalence of asymptomatic malaria infections: a cross-sectional study in rural areas in six departments in Haiti. *Malaria journal.* 2015; 14: 510.
9. Mukry SN, Saud M, Sufaida G, Shaikh K, Naz A, Shamsi TS. Laboratory Diagnosis of Malaria: Comparison of Manual and Automated Diagnostic Tests. *Can J Infect Dis Med Microbiol.* 2017.
10. Joanny F, Lohr SJ, Engleitner T, Lell B, Mordmuller B. Limit of blank and limit of detection of *Plasmodium falciparum* thick blood smear microscopy in a routine setting in Central Africa. *Malar J.* 2014; 13: 234.
11. Boyce MR, O'Meara WP. Use of malaria RDTs in various health contexts across sub-Saharan Africa: a systematic review. *BMC public health.* 2017; 17: 470.
12. Kim S, Nhem S, Dourmg D, Menard D. Malaria rapid diagnostic test as point-of-care test: study protocol for evaluating the VIKIA Malaria Ag Pf/Pan. *Malar J.* 2015; 14: 114.
13. Ashley EA, Touabi M, Ahrer M, Hutagalung R, Htun K, Luchavez J, et al. Evaluation of three parasite lactate dehydrogenase-based rapid diagnostic tests for the diagnosis of falciparum and vivax malaria. *Malar J.* 2009; 8: 241.
14. Maltha J, Gamboa D, Bendezu J, Sanchez L, Cnops L, Gillet P, et al. Rapid diagnostic tests for malaria diagnosis in the Peruvian Amazon: impact of pfrp2 gene deletions and cross-reactions. *PLoS one.* 2012; 7.
15. Gamboa D, Ho MF, Bendezu J, Torres K, Chiodini PL, Barnwell JW, et al. A large proportion of *P. falciparum* isolates in the Amazon region of Peru lack pfrp2 and pfrp3: implications for malaria rapid diagnostic tests. *PLoS one.* 2010; 5.
16. Zheng Z, Cheng Z. Advances in Molecular Diagnosis of Malaria. *Adv Clin Chem.* 2017; 80: 155-192.
17. Padley D, Moody AH, Chiodini PL, Saldanha J. Use of a rapid, single-round, multiplex PCR to detect malarial parasites and identify the species present. *Ann Trop Med Parasitol.* 2003; 97: 131-137.
18. Snounou G, Viriyakosol S, Jarra W, Thaithong S, Brown KN. Identification of the four human malaria parasite species in field samples by the polymerase chain reaction and detection of a high prevalence of mixed infections. *Mol Biochem Parasitol.* 1993; 58: 283-292.
19. Lefterova MI, Budvytiene I, Sandlund J, Farnert A, Banaei N. Simple Real-Time PCR and Amplicon Sequencing Method for Identification of *Plasmodium* Species in Human Whole Blood. *J Clin Microbiol.* 2015; 53: 2251-2257.
20. Kimura M, Miyake H, Kim HS, Tanabe M, Arai M, Kawai S, et al. Species-specific PCR detection of malaria parasites by microtiter plate hybridization: clinical study with malaria patients. *J Clin Microbiol.* 1995; 33: 2342-2346.
21. Ghayour Najafabadi Z, Oormazdi H, Akhlaghi L, Meamar AR, Nateghpour M, Farivar L, et al. Detection of *Plasmodium vivax* and *Plasmodium falciparum* DNA in human saliva and urine: loop-mediated isothermal amplification for malaria diagnosis. *Acta Trop.* 2014; 136: 44-49.
22. Ghayour Najafabadi Z, Oormazdi H, Akhlaghi L, Meamar AR, Raeisi A, Rampisheh Z, et al. Mitochondrial PCR-based malaria detection in saliva and urine of symptomatic patients. *Trans R Soc Trop Med Hyg.* 2014; 108: 358-362.
23. Singh R, Singh DP, Gupta R, Savargaonkar D, Singh OP, Nanda N, et al. Comparison of three PCR-based assays for the non-invasive diagnosis of malaria: detection of *Plasmodium* parasites in blood and saliva. *Eur J Clin Microbiol Infect Dis.* 2014; 33: 1631-1639.
24. Lee MA, Tan CH, Aw LT, Tang CS, Singh M, Lee SH, et al. Real-time fluorescence-based PCR for detection of malaria parasites. *J Clin Microbiol.* 2002; 40: 4343-4345.
25. Hermsen CC, Telgt DS, Linders EH, van de Locht LA, Eling WM, Mensink EJ, et al. Detection of *Plasmodium falciparum* malaria parasites in vivo by real-time quantitative PCR. *Mol Biochem Parasitol.* 2001; 118: 247-251.
26. Notomi T. [Loop-mediated isothermal amplification]. *Nihon rinsho Japanese journal of clinical medicine.* 2007; 65: 957-961.
27. Notomi T, Mori Y, Tomita N, Kanda H. Loop-mediated isothermal amplification (LAMP): principle, features, and future prospects. *J Microbiol.* 2015; 53: 1-5.
28. Han ET, Watanabe R, Sattabongkot J, Khuntirat B, Sirichaisinthop J, Iriko H, et al. Detection of four *Plasmodium* species by genus- and species-specific loop-mediated isothermal amplification for clinical diagnosis. *J Clin Microbiol.* 2007; 45: 2521-2528.
29. Imai K, Tarumoto N, Misawa K, Runtuwene LR, Sakai J, Hayashida K, et al. A novel diagnostic method for malaria using loop-mediated isothermal amplification (LAMP) and MinION nanopore sequencer. *BMC infectious diseases.* 2017; 17: 621.
30. Piera KA, Aziz A, William T, Bell D, Gonzalez IJ, Barber BE, et al. Detection of *Plasmodium knowlesi*, *Plasmodium falciparum* and *Plasmodium vivax* using loop-mediated isothermal amplification (LAMP) in a co-endemic area in Malaysia. *Malar J.* 2017; 16: 29.
31. Singh R, Singh DP, Savargaonkar D, Singh OP, Bhatt RM, Valecha N. Evaluation of SYBR green I based visual loop-mediated isothermal amplification (LAMP) assay for genus and species-specific diagnosis of malaria in *P. vivax* and *P. falciparum* endemic regions. *J Vector Borne Dis.* 2017; 54: 54-60.
32. Tesouro C, Juul S, Arno B, Nielsen CJ, Fiorani P, Frohlich RF, et al. Specific detection of topoisomerase I from the malaria causing *P.*

- falciparum parasite using isothermal rolling circle amplification. Conf Proc IEEE Eng Med Biol Soc. 2012; 2012: 2416-2419.
33. Juul S, Nielsen CJ, Labouriau R, Roy A, Tesauro C, Jensen PW, et al. Droplet microfluidics platform for highly sensitive and quantitative detection of malaria-causing Plasmodium parasites based on enzyme activity measurement. ACS Nano. 2012; 6: 10676-10683.
34. Champoux JJ. DNA topoisomerases: structure, function, and mechanism. Annu Rev Biochem. 2001; 70: 369-413.

**Cite this article**

Hede MS, Knudsen BR (2017) The Use of Highly Sensitive Detection Methods for Eradication of Plasmodium. *Ann Clin Cytol Pathol* 3(8): 1084.